

# Manual Supplement

Manual Title:	150X Series Calibration Supplement	Issue:	<b>2</b>
Part Number:	2465477	Issue Date:	2/09
Print Date:	November 2005	Page Count:	1
Revision/Date:	1, 7/07		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1

On page 18, Table 5, **Step 8**, under **Display Reading Limits** for **1507/1508**:

Change: 0.6 to 0.15

To: 0.05 to 0.15

## Change #2

On page 25, Table 10, **Adjustment Step C-03**, under **Input Value** Column:

Change: 0.18 mA 0 Hz

To: 20  $\mu$ A, 0 Hz