

Manual Supplement

Manual Title:	33X Calibration	Supplement Issue:	3
Part Number:	1618765	Issue Date:	4/10
Print Date:	June 2001	Page Count:	2
Revision/Date:	3, 4/06		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 8, under **5520A Output**, $\tilde{\text{A}}$ Amps AC,

Change:

10 A 50 Hz (333)

To:

8.0 A 50 Hz (333)

Under **5520A Output**, *INRUSH,

Change:

*INRUSH	**5 A 60 Hz
---------	-------------

To:

*INRUSH	**0.1 A 60 Hz
---------	---------------

On Page 9, under **5520A Output**, **Lower Limit** and **Upper Limit**, $\tilde{\text{A}}$ Amps AC, delete the entire row:

10 A 50 Hz	187.5 A AC	212.5 A AC
------------	------------	------------

Under **5520A Output** *INRUSH and Frequency (337),

Change:

*INRUSH	**5 A 60 Hz
---------	-------------

To:

*INRUSH	**0.1 A 60 Hz
---------	---------------

Change:

Frequency (337)	5 A 10 Hz
	10 A 300 Hz

To:

Frequency (337)	0.1 A 10 Hz
	0.2 A 300 Hz

Change #2

On page 5, under **Performance Test**, at the end of the first paragraph, add the following:

Performance tests should be done annually (every 12 months) to ensure that the Meter is within its accuracy specifications.

Change #3, 54734

On page 3, under Range for the **333**:

Change: 0 - 400.00 A

To: 0 - 400.0 A