

# Manual Supplement

Manual Title:	32X Calibration	Supplement Issue:	4
Part Number:	1631636	Issue Date:	12/09
Print Date:	October 2001	Page Count:	2
Revision/Date:	1, 12/03		

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1

On page 9, Table 3, replace the **DC Volts** section, with the following:

Functional Test	5520A Output	Nominal	Low Limit	High Limit
<b>DC Volts</b>	-350 V	-350.0 V	-354.0	-346.0
	35 V	35.0 V	34.1	35.9
	350 V	350.0 V	346.0	354.0
	600 V	600 V	589	611

## Change #2 w2030084

On page 2, replace the TUV symbol with the following:



## Change #3

On page 8, Table 2, replace the **AC Volts**, **Low Limit** and **High Limit** sections with the following:

Low Limit	High Limit
34.1	35.9
345.3	354.7
588	612
34.1	35.9
345.3	354.7
588	612

On page 9, Table 3, replace the **AC Volts**, **Low Limit** and **High Limit** sections with the following:

Low Limit	High Limit
34.1	35.9
345.3	354.7
588	612
34.1	35.9
345.3	354.7
588	612

## Change #4

On page 5, **Table 1**, under **Equipment**, replace the following:

From: AC Calibrator

To: Multifunction Calibrator