Manual Supplement

Manual Title:360 CalibrationSupplement Issue:1Part Number:Web OnlyIssue Date:6/08Print Date:October 2007Page Count:1

Revision/Date:

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

- 1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
- 2. No revision letter is indicated at the beginning of the change.



360 Calibration Manual Supplement

Change #1, 47120

On page 3, under *Specifications*, replace the **Temperature Coefficient**, section with the following:

Temperature Coefficient
<18 °C or >28 °C

Measuring range: 0 to 50 A ± 0.08 % of reading/°C + 0.5 digit Measuring range: 50 to 60 A ± 0.3 % of reading/°C ± 0.5 digit

On page 4, under *Electrical Specifications*, add:

Frequency: 50 Hz ±1 Hz, 60 Hz ±1 Hz

6/08